



DECLARATION OF CONFORMITY
Regarding Medical Device Regulation (EU) 2017/745



Manufacturer: Sichuan Ai Doctor Medical Technology Co., Ltd.

Address: 333 Yongke Road, Yongsheng Town, Wenjiang District, Chengdu City,
Sichuan Province, China

EC Representative: SUNGO Europe B.V.

Address: Olympisch Stadion 24, 1076DE Amsterdam, Netherlands

Product Name: Medical Face Mask **HYCISUN[®]**

SRN: _____ / _____

Basic UDI-DI: _____ / _____

Classification Class I

Rule: Rule 5, Annex VIII, Regulation (EU) 2017/745

Conformity Assessment Procedure: Annex II+III of Regulation (EU) 2017/745

We herewith declare that the above-mentioned products meet the requirements of Medical Device Regulation (EU) 2017/745 and the following harmonized standards.

EN ISO 14971: 2012

ISO 10993-1: 2018

EN ISO 15223-1: 2016

EN ISO 10993-5: 2009

EN 1041:2008+A1:2013

EN ISO 10993-10: 2013

EN14683:2019+AC:2019



On behalf of SUNGO Europe office, I confirmed we are EU REP of the company who issue this document.

Signature: Yingbo Tang

Name / Position: Yingbo Tang / General Manager

Date: 2020.11.23

Place: China



[Handwritten Signature]





Authorized Signature (S)



*On behalf of SUNGO Europe office
EU REP of the company who issue*

Author

Annex

Product	Model	Basic UDI-DI	Picture
Medical Face Mask	Earloop Type 170*95MM		
Medical Face Mask	Tie-on Type 170*95MM		
Medical Face Mask	HS0402R (Black/Schwarz)		
<p><i>On behalf of SUNGO Europe office, I confirmed we are EU REP of the company who issue this document.</i></p>			<p><i>I confirmed we are this document.</i></p>
 <p><i>[Signature]</i> Authorized Signature (S)</p>			<p><i>[Signature]</i> Authorized Signature (S)</p>

Signature: Yingbo Tang
 Name / Position: Yingbo Tang / General Manager

Date: 2020.11.23
 Place: China



DECLARATION OF CONFORMITY
Regarding Medical Device Regulation (EU) 2017/745



Manufacturer: Sichuan Ai Doctor Medical Technology Co., Ltd.

Address: 333 Yongke Road, Yongsheng Town, Wenjiang District, Chengdu City,
Sichuan Province, China

EC Representative: SUNGO Europe B.V.

Address: Olympisch Stadion 24, 1076DE Amsterdam, Netherlands

Product Name: Surgical Mask **HYGISUN[®]**

SRN: _____ / _____

Basic UDI-DI: _____ / _____

Classification Class I

Rule: Rule 5, Annex VIII, Regulation (EU) 2017/745

Conformity Assessment Procedure: Annex II+III of Regulation (EU) 2017/745

We herewith declare that the above-mentioned products meet the requirements of *On behalf of SUNGO Europe*
Medical Device Regulation (EU) 2017/745 and the following harmonized standards. *EU REP of the company who*

EN ISO 14971: 2012

ISO 10993-1: 2018

EN ISO 15223-1: 2016

EN ISO 10993-5: 2009

EN 1041:2008+A1:2013

EN ISO 10993-10: 2013

EN14683:2019+AC:2019

*On behalf of SUNGO Europe office, I confirmed we are
EU REP of the company who issue this document.*


Signature: Yingbo Tang
Name / Position: Yingbo Tang / General Manager









Date: 2020.11.23

Authorized Signature (S)

Place: China



Annex

Product	Model	Basic UDI-DI	Picture
Surgical Mask	Earloop Type 170*95MM Non-sterile		
Surgical Mask	Tie-on Type 170*95MM Non-sterile		
Surgical Mask	HS0402R (Black/Schwarz) Non-sterile		
<p><i>On behalf of SUNGO Europe office, I confirmed we are EU REP of the company who issue this document.</i></p>			<p><i>Office, I confirmed we are sue this document.</i></p>
			
		<p><i>Authorized Signature (S)</i></p>	<p><i>Authorized Signature (S)</i></p>

Signature: Yingbo Tang
 Name / Position: Yingbo Tang / General Manager

Date: 2020.11.23
 Place: China





> Retouradres Postbus 16114 2500 BC Den Haag

SUNGO Europe B.V.
T.a.v. de heer R. Luo
Olympisch Stadion 24
1076 DE Amsterdam

Datum: 15 december 2020
Betreft: notificatie medische hulpmiddelen klasse I

Geachte heer Luo,

Hierbij bevestig ik de ontvangst op 1 december 2020 van de notificatie van de medische hulpmiddelen klasse I, die bedrijf Sichuan Ai Doctor Medical Technology Co., Ltd., met Europees gemachtigde SUNGO Europe B.V. , als fabrikant overeenkomstig Verordening (EU) 2017/745 (MDR) op de markt wenst te gaan brengen.

De producten zijn onder volgend kenmerk geregistreerd. Ik verzoek u om in alle verdere correspondentie over een of meer van deze producten het bijbehorende kenmerk te vermelden en het bij telefoongesprekken bij de hand te houden.

Medical Face Mask
(geen merknaam) (NL-CA002-2020-54533)
Surgical Mask
(geen merknaam) (NL-CA002-2020-54534)

Ik wijs u erop dat medische hulpmiddelen die op de markt gebracht worden volgens de MDR over een systeem voor hulpmiddelindicatie (UDI) moeten beschikken¹ en dat fabrikanten, gemachtigden en importeurs in de Europese databank voor Europese hulpmiddelen (Eudamed) moeten worden geregistreerd². Bijlage VI van de MDR bevat de bij de registratie te verstrekken gegevens.

Op dit moment is Eudamed nog niet in gebruik, zodat het wat betreft het bovenstaande voldoende is dat u uw producten overeenkomstig de huidige wet- en regelgeving hebt genotificeerd.

Zodra Eudamed volledig in gebruik is, wordt de fabrikant of diens gemachtigde geacht binnen achttien maanden bovenstaande hulpmiddelen te registreren in Eudamed.³

¹ O.g.v. art. 29 MDR.

² O.g.v. art. 31 MDR.

³ www.camd-europe.eu/wp-content/uploads/2018/05/FAQ_MDR_180117_V1.0-1.pdf. Zie vraag en antwoord nummer 20.

Farmatec

Bezoekadres:
Hoftoren
Rijnstraat 50
2515 XP Den Haag
T 070 340 6161

<http://hulpmiddelen.farmatec.nl>

Inlichtingen bij:

M.P. Meijer - Michiels

medische_hulpmiddelen@
minvws.nl

Ons kenmerk:

CIBG-20205750

Bijlagen

-

Uw aanvraag


1 december 2020

*Correspondentie uitsluitend
richten aan het retouradres met
vermelding van de datum en het
kenmerk van deze brief.*

Tevens wijs ik u er voor de goede orde nog op dat de registratie van uw mededeling betreffende de aflevering van de bovengenoemde producten slechts een administratieve handeling betreft. Deze ontvangstbevestiging behelst dan ook geen besluit betreffende de kwalificatie van de desbetreffende producten als medisch hulpmiddel in de zin van art. 1 WMH, noch betreffende de indeling in risicoklasse I.

De Minister voor Medische Zorg en Sport,
namens deze,

Afdelingshoofd
Farmatec

A handwritten signature in blue ink, appearing to read 'M.J. van de Velde', written over a horizontal line.

Dr. M.J. van de Velde



中国认可
国际互认
检测
TESTING
CNAS L0412

检测报告

(Test Report)

No. GOLCXHJP953425L1

样品名称 (Sample Description)	一次性医用口罩（非无菌） Medical Face Mask (non-sterile)
委托单位 (Applicant)	四川艾医生医疗科技有限公司 Sichuan Ai Doctor Medical Technology Co.,Ltd.

声 明
Statement

1. 本报告无检验检测专用章、报告骑缝章和批准人签章无效。
This report is invalid without special seal of inspection, cross-page seal and the approver's signatures.
2. 本报告页面所使用“PONY”、“谱尼”字样为本单位的注册商标,其受《中华人民共和国商标法》保护,任何未经本单位授权的擅自使用和仿冒、伪造、变造“PONY”、“谱尼”商标均为违法侵权行为,本单位将依法追究其法律责任。
The pattern and characters of "PONY" and "谱尼" used in this report are protected by the trademark law of the People's Republic of China. Any unauthorized usage, counterfeit, forgery and alteration of trademarks of "PONY" and "谱尼" are the violations of the law. The PONY has the right to pursue all legal liabilities of the subject of the delict.
3. 委托单位对报告数据如有异议,请于报告完成之日起十五日内(初级农产品报告请于报告收到之日起五日内)向本单位书面提出复测申请,同时附上报告原件并预付复测费。
If the applicant has any questions about the results, shall provide a written retest application with the original report, and prepay the retest fees to PONY within fifteen days since the approval date (as an exception, it shall be within five days since the date received for the primary agriculture products report).
4. 委托单位办理完毕以上手续后,本单位会尽快安排复测。如果复测结果与异议内容相符,本单位将退还委托单位的复测费。
After the applicant finishes the procedure mentioned above, PONY shall arrange the retest as soon as possible. If the retest result accords with the applicant dissent, PONY shall refund the retest fees.
5. 不可重复性或不能进行复测的实验,不进行复测,委托单位放弃异议权利。
Tests that can not be repeated and tested shall not be carried out again.
6. 委托单位对样品的代表性和资料的真实性负责,否则本单位不承担任何相关责任。
The applicant should undertake the responsibility for the provided samples' representativeness and document authenticity. Otherwise, PONY has not any relevant responsibilities.
7. 本报告仅对所测样品负责,报告数据仅反映对所测样品的评价,对于报告及所载内容的使用、使用所产生的直接或间接损失及一切法律后果,本单位不承担任何经济和法律后果。
This report is only responsible for the provided sample. The test results only represent the evaluation of the tested sample. PONY will not be responsible for any economical or legal liability generated from direct or indirect usage of the test report.
8. 本单位有权在完成报告后按规定方式处理所测样品。
PONY has the right to dispose the tested sample by rules, after approval of the test report.
9. 本单位保证工作的客观公正性,对委托单位的商业信息、技术文件等商业秘密履行保密义务。
PONY assures objectivity and impartiality of the test, and fulfills the obligation of confidentiality for applicant's commercial information, and technique document.
10. 本报告私自转让、盗用、冒用、涂改、未经本单位批准的复制(全文复制除外)或以其它任何形式的篡改均属无效,本单位将对上述行为追究其相应的法律责任。
The report is invalid in case of illegal transfer, embezzlement, imposture, modification or any altering, reproducing except in full, without approval of PONY. PONY shall investigate and affix the applicant's legal liability accordingly.

▲ 防伪说明 (Anti-counterfeiting Description):

- (1) 报告编号是唯一的;
The test report has exclusive report code.
- (2) 报告采用特制防伪纸张印制,纸张表面带有"PONY"防伪纹路,该防伪纹路不支持复印,即复制件不会带有"PONY"防伪纹路。
The test report is printed by anti-copying paper whose surface shows "PONY" security print with specific anticounterfeiting technique. Security print will disappear after copying. Duplicates are not expected to give "PONY" security print under any circumstances.



全国服务热线
400-819-5688

WWW.PONYTEST.COM

扫描二维码

关注谱尼测试微信

公众号 PONY4008195688



北京实验室: (010) 83055000	武汉实验室: (027) 83997127	哈尔滨实验室: (0451)58627755
上海实验室: (021) 64851999	长春实验室: (0431)85150908	石家庄实验室: (0311)85376660
青岛实验室: (0532)88706866	大连实验室: (0411)87336618	乌鲁木齐实验室: (0991) 6684186
深圳实验室: (0755)26050909	郑州实验室: (0371)69350670	呼和浩特实验室: (0471) 3450025
天津实验室: (022) 23607888	西安实验室: (029) 89608785	杭州实验室: (0571)85806807
苏州实验室: (0512)62997900	太原实验室: (0351) 7555762	宁波实验室: (0574)87977185
		温州实验室: (0577)88271060
		合肥实验室: (0551)63843474
		广州实验室: (020) 89224310
		厦门实验室: (0592)5568048
		成都实验室: (028) 87702708

检测结果

(Test Results)

No. GOLCXHJP953425L1

第 1 页, 共 3 页 (page 1 of 3)

样品名称 (Sample Description)	一次性医用口罩 (非无菌) Medical Face Mask (non-sterile)	样品规格 (Sample Specification)	170mm*95mm
委托单位 (Applicant)	四川艾医生医疗科技有限公司 Sichuan Ai Doctor Medical Technology Co.,Ltd.	商标 (Trade Mark)	—
到样日期 (Received Date)	2020-11-27	生产日期或批号 (Manufacturing Date or Lot No.)	2020.11.01 20201101
检测日期 (Test Date)	2020-11-27~2020-12-07	样品等级 (Sample Grade)	—
样品状态 (Sample Status)	正常 Normal	检测类别 (Test Type)	委托检测 Commissioning Test
检测项目 (Test Items)	见下页 See next page	检测环境 (Test Environment)	符合要求 To meet the requirements
检测方法 (Test Methods)	见下页 See next page		
所用主要仪器 (Main Instruments)	口罩细菌过滤效率检测仪 等 Mask bacteria filtration efficiency detector etc.		
备注 (Note)	1.型号: Earloop Type Model: Earloop Type 2.生产单位/受检单位: 四川艾医生医疗科技有限公司 Manufacturer/Tested company: Sichuan Ai Doctor Medical Technology Co.,Ltd. 3.以上样品信息由委托单位提供 The information of sample was provided by the applicant 4.该报告中检测方法由委托单位指定。 The testing methods mentioned in this report were designated by the applicant. 5.限值标准: BS EN 14683:2019 (IIR 型) Limit Standard: BS EN 14683:2019(Type IIR)		
	编制人 (Edited by)	王会兴	
	审核人 (Checked by)	王明	
	批准人 (Approved by)	孙兆增	
	签发日期 (Issued Date)	2020 年 12 月 07 日	

检测结果 (Test Results)

No. GOLCXHJP953425L1

第 2 页, 共 3 页 (page 2 of 3)

序号 (S/N)	检测项目 (Test Item)	单位 (Unit)	限值 (Limit)	检测结果 (Test Result)			单项结论 (Evaluation)	检测方法 (Test Method)
1	细菌过滤效率 (BFE) Bacterial filtration efficiency(BFE)	%	≥98	99.82			符合 Pass	BS EN 14683:2019 附录 B Appendix B
				99.91				
				99.87				
				99.91				
				99.78				
2	压力差 Differential pressure	Pa/cm ²	<60	A	B	C	符合 Pass	BS EN 14683:2019 附录 C Appendix C
				1-1	31.9	29.6		
				1-2	26.9			
				1-3	25.9			
				1-4	29.8			
				1-5	33.4			
				2-1	28.8	29.3		
				2-2	27.8			
				2-3	25.7			
				2-4	31.0			
				2-5	33.3			
				3-1	30.3	28.3		
				3-2	28.7			
				3-3	25.6			
				3-4	28.5			
				3-5	28.2			
				4-1	35.3	30.9		
				4-2	26.3			
				4-3	28.6			
				4-4	31.9			
				4-5	32.3			
				5-1	29.0	28.6		
				5-2	25.7			
				5-3	22.7			
				5-4	34.2			
5-5	31.4							

☎ Hotline 400-819-5688

www.ponytest.com

PONY-BG186-01A-1-B-1-01-2019A

谱尼测试集团股份有限公司 Pony Testing International Group Co. Ltd.

公司地址: 北京市海淀区锦带路 66 号院 1 号楼 5 层 101
Address: 101, 5F, Building 1, No.66 Jindai Road, Haidian District, Beijing, China

检测地址: 北京市海淀区锦带路 66 号院 1 号楼
Testing address: Building 1, No.66 Jindai Road, Haidian District, Beijing, China

电话: 010-83055000 传真: 010-82619629
Tel: 010-83055000 Fax: 010-82619629

检测结果 (Test Results)


No. GOLCXHJP953425L1

第 3 页, 共 3 页 (page 3 of 3)

序号 (S/N)	检测项目 (Test Item)	单位 (Unit)	限值 (Limit)	检测结果 (Test Result)	单项结论 (Evaluation)	检测方法 (Test Method)
3	抗溅压力 Splash resistance pressure	kPa	≥16.0	32 个试样均 > 16.0 Splash resistance pressure of 32 samples were all greater than 16.0	符合 Pass	ISO 22609:2004
4	微生物洁净度 Microbial cleanliness	cfu/g	≤30	23	符合 Pass	BS EN 14683:2019 附录 D Appendix D
				17		
				16		
				17		
				17		

备注 Note: A-试样编号-测试区域编号 Test Specimen number-Test area number; B-每个测试区域的压力差 Differential pressure for each test area; C-每个试样的平均压差 The averaged differential pressure for each test specimen.

照片 Photo:



——以下空白——
(End of Report)



SUBJECT Physical & Microbiological Test

TEST LOCATION TÜV SÜD China

TÜV SÜD Products Testing (Shanghai) Co., Ltd.
B-3/4, No.1999 Du Hui Road, Minhang District
Shanghai 201108, P.R. China

CLIENT NAME Sichuan Ai Doctor Medical Technology Co.,Ltd.

CLIENT ADDRESS No.333, Yongke Road, Yongsheng Town, Chengdu Cross-strait Science And
Technology Industrial Development Park, Wenjiang District, Chengdu, Sichuan,
China

TEST PERIOD 26-Apr-2020~08-May-2020

Prepared By

Bella Xu

(Bella Xu)
Report Drafter

Authorized By



(Leo Liu)
Authorized Signatory

Note: (1) General Terms & Conditions as mentioned overleaf. (2) The results relate only to the items tested.(3) The test report shall not be reproduced except in full without the written approval of the laboratory.(4) Without the agreement of the laboratory, the client is not authorized to use the test results for unapproved propaganda.

Chemical/Microbiology Laboratory:
TÜV SÜD Products Testing (Shanghai) Co.,
Ltd.
B-3/4, No.1999 Du Hui Road, Minhang District
Shanghai
201108
P.R. China

Phone : +86 (21) 6037 6375
Fax : +86 (21) 6037 6345
Email: food.chem@tuv-sud.cn
Webpage: www.tuv-sud.cn

Regional Head Office:
TÜV SÜD Certification and Testing
(China) Co., Ltd.
No.151 Heng Tong Road Shanghai
200 070 P.R.China

TUV®

TEST REPORT

Sample Description : Surgical Mask
Sample Quantity : 50 pieces
Lot Number/Batch Code : 20200401
Specification : Earloop Type
Size : /
Brand Name : /

Remark: The above information was provided by applicant.

Summary of Test Results

No.	Test Item	Test Method	Test Standard Type II R	Judgement
1	Bacterial Filtration Efficiency Test (BFE), %	EN 14683:2019+AC:2019(E) Annex B	≥ 98	Pass
2	Differential Pressure Test (Pa/cm ²)	EN 14683:2019+AC:2019(E) Annex C	< 60	Pass
3	Synthetic Blood Penetration Test (kPa)	ISO 22609:2004	≥ 16.0	Pass
4	Microbial Cleanliness Test (CFU/g)	EN 14683:2019+AC:2019(E) Annex D	≤ 30	Pass

Note: Pass = Meet customer requirements;
Fail = Fail customer requirements;
= No comment;
N.D. = Not detected.

Photo of Samples



Results

No.	Test Item	Test Result
1	Bacterial Filtration Efficiency (BFE) Test	Specimen 1#: 98.2% Specimen 2#: 98.0% Specimen 3#: 98.1% Specimen 4#: 98.1% Specimen 5#: 98.1%
2	Differential Pressure Test	26.6 Pa/cm ²
3	Synthetic Blood Penetration Test	Specimen 1#~32#: None seen
4	Microbial Cleanliness Test	Specimen 1#: <1 CFU/g Specimen 2#: 1 CFU/g Specimen 3#: <1 CFU/g Specimen 4#: <1 CFU/g Specimen 5#: <1 CFU/g

Bacterial Filtration Efficiency (BFE) Test

1. Purpose

For evaluating the bacterial filtration efficiency (BFE) of mask.

2. Sample description was given by client

Sample description : Surgical Mask
Specification : Earloop Type
Lot Number : 20200401
Sample Receiving Date : 2020-04-26

3. Test Method

EN 14683:2019+AC:2019(E) Annex B

4. Apparatus and materials

- 4.1 *Staphylococcus aureus* ATCC 6538.
- 4.2 Peptone water.
- 4.3 Tryptic Soy Broth(TSB).
- 4.4 Tryptic Soy Agar(TSA).
- 4.5 Bacterial filtration efficiency test apparatus.
- 4.6 Six-stage viable particle Anderson sampler.
- 4.7 Flow meters.

5. Test specimen

- 5.1 As requested by client, take a total of 5 test specimens.
- 5.2 Prior to testing, condition all test specimens for a minimum of 4 h at (21±5)°C and (85±5)% relative humidity.

6. Procedure

- 6.1 Preparation of the bacterial challenge: Dilute the culture in peptone water to achieve a concentration of approximately 5×10^5 CFU/mL.
- 6.2 Adjust the flow rate through the Anderson sampler to 28.3 L/min.
- 6.3 Deliver the challenge to the nebulizer using a syringe pump. Purge tubing and nebulizer of air bubbles.
- 6.4 Perform a positive control run without a test specimen to determine the number of viable aerosol particles being generated. The mean particle size (MPS) of the aerosol will also be calculated from the results of these positive control plates.
 - 6.4.1 Initiate the aerosol challenge by turning on the air pressure and pump connected to the nebulizer. Immediately begin sampling the aerosol using the Anderson sampler.
 - 6.4.2 Time the challenge suspension to be delivered to the nebulizer for 1 min.
 - 6.4.3 Time the air pressure and Anderson sampler to run for 2 min.
 - 6.4.4 At the conclusion of the positive control run, remove plates from the Anderson sampler.
- 6.5 Place new agar plates into Anderson sampler and clamp the test specimen into the top of the Anderson sampler, with the inside of the specimen facing towards the bacterial challenge (test area: 77cm²).
- 6.6 Repeat the challenge procedure for each test specimen.
- 6.7 Repeat a positive control after completion of the sample set.
- 6.8 Perform a negative control run by collecting a 2 min sample of air from the aerosol chamber. No bacterial challenge should be pumped into the nebulizer during the collection of the negative control.
- 6.9 Incubate agar plates at $(37 \pm 2)^\circ\text{C}$ for (20 to 52) h.
- 6.10 Count each of the six-stage plates of the Anderson sampler.

7. Calculation

Total the count from each of the six plates for the test specimens and positive controls, as specified by the manufacture of Anderson sampler. The filtration efficiency percentages are calculated as follows:

$$\text{BFE} = (C - T) / C \times 100$$

T is the total plate count for the test specimen.

C is the mean of the total plate counts for the two positive controls.



8. Test results*

<i>P</i> Value Stage Number	Positive Control (A)	Positive Control (B)	Negative Control	Specimen 1#	Specimen 2#	Specimen 3#	Specimen 4#	Specimen 5#
1	53	98	0	0	0	0	1	1
2	112	136	0	2	1	1	2	1
3	130	165	0	1	1	2	6	1
4	213	298	0	3	1	6	7	6
5	1518	1518	0	26	29	27	27	29
6	434	475	0	16	19	13	7	12
Total (<i>T</i>), CFU	2460	2690	<1	48	51	49	50	50
Average (<i>C</i>), CFU	$2.6 \times 10^3 = (P_A + P_B) / 2$							
BFE, %				98.2	98.0	98.1	98.1	98.1
Requirements	≥ 98							
Remarks	<i>P</i> is the value of corresponding corrected particle counts as specified by the manufacturer of the cascade impactor. <i>T</i> is the total of <i>P</i> value for the test specimen. <i>C</i> is the mean of the total of <i>P</i> value of the two positive controls.							

Differential pressure Test

1. Purpose

The purpose of the test was to measure the differential pressure of masks.

2. Sample description was given by client

Sample description : Surgical Mask
Specification : Earloop Type
Lot Number : 20200401
Sample Receiving Date : 2020-04-26

3. Test Method

EN 14683:2019+AC:2019(E) Annex C

4. Apparatus and materials

Differential pressure testing instrument

5. Test specimen

- 5.1 Test specimen are complete masks or shall be cut from masks. Each specimen shall be able to provide 5 different circular test areas of 2.5 cm in diameter.
5.2 Prior to testing, condition all test specimens for a minimum of 4 h at (21±5) °C and (85±5)% relative humidity.

6. Procedure

- 6.1 Without a specimen in place, the holder is closed and the differential manometer is zeroed. The pump is started and the flow of air adjusted to 8 L/min.
6.2 The pretreated specimen is placed across the orifice (total area 4.9cm², test area diameter 25mm) and clamped into place so as to minimize air leaks.
6.3 Due to the presence of an alignment system the tested area of the specimen should be perfectly in line and across the flow of air.
6.4 The differential pressure is read directly.
6.5 The procedure described in steps 6.1-6.4 is carried out on 5 different areas of the mask and readings averaged.

Results:

Specimen	Test Results* (Pa/cm ²)	Average (Pa/cm ²)	Requirements	Judgement
1#	27.4	26.6	< 60	Pass
2#	29.1			
3#	22.6			
4#	27.5			
5#	26.5			

TUV
SUD
告
1
*

Synthetic Blood Penetration Test

1. Purpose

For evaluation of resistance of masks to penetration by a fixed volume of synthetic blood at a high velocity.

2. Sample description was given by client

Sample description : Surgical Mask
Specification : Earloop Type
Lot Number : 20200401
Sample Receiving Date : 2020-04-26

3. Test Method

ISO 22609:2004

4. Apparatus and materials

- 4.1 Synthetic blood.
- 4.2 Tensiometer.
- 4.3 Synthetic blood penetration test apparatus;
- 4.4 Targeting plate.
- 4.5 Air pressure source.
- 4.6 Ruler.
- 4.7 Balance.
- 4.8 Controlled temperature and humidity chamber.

5. Test specimen

- 5.1 As requested by client, take a total of 32 test specimens.
- 5.2 Prior to testing, condition all test specimens for a minimum of 4h at (21±5)°C and (85±5) % relative humidity.

6. Procedure

- 6.1 Prepare the synthetic blood (40~44 mN/m) for the test.
- 6.2 Determine the density of the synthetic blood.
- 6.3 Fill the reservoir with new synthetic blood.
- 6.4 Position the test specimen 30.5 cm (12 in.) from the exit of the canula.
- 6.5 Set the reservoir pressure to the approximate pressure.
- 6.6 Place the targeting plate approximately 1 cm away from the mask.
- 6.7 Set the valve timer to 0.5 s. Collect and weigh the amount of fluid delivered (before the targeting hole).

- 6.8 Set the valve timer to 1.5 s. Collect and weigh the amount of fluid delivered (before the targeting hole).
- 6.9 Calculate the difference in weight of the two spurts. For a test fluid with a density of 1.003, Table 1 gives the target difference in weight plus lower and upper limits for a velocity range within 2% of the target.

Table 1 Target weight difference

Fluid Pressure (mmHg)	Weight difference for 1s difference in spurt duration (g)		
	Min.	Target	Max.
120	3.002	3.063	3.124

- 6.10 Adjust the reservoir pressure and repeat steps 6.7 to 6.9 until the weight difference is within the target range.
- 6.11 Record the weight difference for the spurts exiting the nozzle.
- 6.12 Record the pressure in the reservoir.
- 6.13 Set the valve time to 0.5 s. Collect and weigh the amount of fluid passing through the targeting hole.
- 6.14 Set the valve time to 1.5 s. Collect and weigh the amount of fluid passing through the targeting hole.
- 6.15 The difference in weight between the 0.5 s and 1.5 s spurts through the targeting plate shall be within +2 % ~ -5 % of the difference in weight from the nozzle.
- 6.16 If the differential weight is less than 95 % of the weight difference exiting the nozzle, check the aim of the stream to make sure it is passing cleanly through the targeting hole.
- 6.17 If the differential weight is more than 102 % of the weight difference exiting the nozzle, repeat the weight measurements exiting the nozzle (steps 6.7 to 6.11).
- 6.18 For standard synthetic blood, the timer duration can be estimated using the formula:
(p is the density of the test fluid.) $t = 0.5 + (2 \times p - g \text{ at } 0.5 \text{ s}) / (g \text{ at } 1.5 \text{ s} - g \text{ at } 0.5 \text{ s})$.
- 6.19 Record the timer setting to use as the starting point for subsequent testing.
- 6.20 Mount a test specimen on the specimen holding fixture. If the mask contains pleats, spread the pleats out when mounting the mask onto the fixture to present a single layer of material as the target area.
- 6.21 Squirt the synthetic blood onto the test specimen for the calculated time. Ensure that the synthetic blood hits the target area of mask.
- 6.22 Inspect the inside surface for synthetic blood penetration within 10 s of squirting the synthetic blood against the target area.
- 6.23 Report the results (none / penetration) for each test specimen at the test pressure.



Results:

Specimen	Test Results*	Requirements	Judgement
1#	None Seen	Pass Pressure at 16.0 kPa (120mmHg)	Pass
2#	None Seen		Pass
3#	None Seen		Pass
4#	None Seen		Pass
5#	None Seen		Pass
6#	None Seen		Pass
7#	None Seen		Pass
8#	None Seen		Pass
9#	None Seen		Pass
10#	None Seen		Pass
12#	None Seen		Pass
13#	None Seen		Pass
14#	None Seen		Pass
15#	None Seen		Pass
16#	None Seen		Pass
17#	None Seen		Pass
18#	None Seen		Pass
19#	None Seen		Pass
20#	None Seen		Pass
22#	None Seen		Pass
23#	None Seen		Pass
24#	None Seen		Pass
25#	None Seen		Pass
26#	None Seen		Pass
27#	None Seen		Pass
28#	None Seen		Pass
29#	None Seen		Pass
30#	None Seen		Pass
31#	None Seen		Pass
32#	None Seen		Pass

Microbial Cleanliness Test

1. Purpose

The purpose of the test was to measure microbial cleanliness of mask.

2. Sample description was given by client

Sample description : Surgical Mask
Specification : Earloop Type
Lot Number : 20200401
Sample Receiving Date : 2020-04-26

3. Test Method

According to EN ISO 11737-1:2018 to determine the microbial cleanliness of mask material, and refer to the procedure as described in EN 14683:2019+AC:2019(E) Annex D

4. Apparatus and materials

- 4.1 Orbital shaker.
- 4.2 0.45 um filter.
- 4.3 Tryptic Soy Agar (TSA).
- 4.4 Sabouraud Dextrose Ager (SDA) with chloramphenicol.
- 4.5 Formula of Extraction Liquid: 1g/L peptone, 5g/L NaCl and 2g/L Tween 20.
- 4.6 Extraction apparatus.

5. Test specimen

- 5.1 As requested by client, take a total of 5 mask samples.
- 5.2 Mask samples for testing are provided in the original primary packaging.
- 5.3 Condition at (18 to 26)°C and (45 to 65)% relative humidity during testing.

6. Procedure

- 6.1 Five test specimens are selected from the top, bottom and 3 randomly chosen marks.
- 6.2 The mask is aseptically removed from the packaging and placed in a sterile 500 mL bottle containing 300 mL of extraction liquid.
- 6.3 The bottle is laid down on an orbital shaker and shaken for 5 min at 250 rpm.
- 6.4 After extracting, 100mL of the extraction liquid is filtered through a 0.45 um filter and laid down on a TSA plate for the total viable aerobic microbial count. Another 100 mL aliquot of the same extraction liquid is filtered in the same way and the filter plated on SDA for fungi enumeration.
- 6.5 The plates are incubated for 3 days at 30°C and 7 days at (20 to 25)°C for TSA and SDA plates respectively.
- 6.6 Calculate the colonies of each agar plate.

7. Calculation

For each test specimen calculate the microbial cleanliness as follows by counting the total colonies of the TSA and SDA plates.



Results*:

Specimen	Colonies of the TSA Plate	Colonies of the SDA Plate	Microbial Cleanliness, (CFU/g)	Requirements	Judgement
1#	0	0	<1	EN14683:2019+AC:2019(E) Annex D EN ISO 11737-1:2018 ≤ 30 CFU/g	Pass
2#	0	1	1		
3#	0	0	<1		
4#	0	0	<1		
5#	0	0	<1		

Note:

- 1.*denotes this test was carried out by external laboratory assessed as competent.
- 2.This report is for internal use only such as internal scientific research ,education, quality control, product R&D.

-END OF THE TEST REPORT-

